Certificate of Calibation

Coti Global Sensors
5709 Hwy 53
Harvest,AL 35749

Calibration Date	10/24/2011
Calibration Renewal	10/24/2012

Model	CG-26S5		Standard	CP26S5
Capacity	300K		Capacity	500K
Serial Number	2011027002		Serial Number	C11396
			NSWCML	FX9092
Indicator	GSE 250			
Serial Number	3259526			

Calibration

Weight Applied			Indicator	
0			0	
150K			150250	
300K			300000	
150K			150000	
0			0	

Inspector	Donald Hawkins
Date	10/24/2011