



# TEXAS DEPARTMENT OF AGRICULTURE

COMMISSIONER SID MILLER

Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942

Phone: (979) 542-3231 - Fax: (888) 205-7741

Test Number

G-000006283

## CALIBRATION CERTIFICATE

FOR  
3 - Weight Kits

SUBMITTED BY  
Bastrop Scale Company  
PO Drawer 2100  
Bastrop, Texas 78602

The standards and measurements of the Texas Department of Agriculture, Giddings Metrology Laboratory, are traceable to the SI and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of the uncertainty reported by this laboratory. The data below applies only to the artifacts identified in this report at the time of test.

Test Date: 09/13/2017  
Calibration Due: 09/30/2018

Received Date: 9/12/2017  
Condition Received: Acceptable

Temperature Range: 18 °C to 27 °C  
Relative Humidity Range: 40 % to 60 %

Procedure: NISTIR 6969, SOP No. 8, Modified Substitution

Balances: CCE5003-SN:21411465, CCE2004-SN:28601871, MSA225S-SN:28512376, CCE36-SN:28601985

Mass Standards: Giddings Metrology Laboratory Mass Echelon III Standards

The expanded standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, and a component of uncertainty to account for any observed deviations that have a significant effect on the calibration. No component is included in the expanded uncertainty for the effects of magnetism. The expanded uncertainty given is in compliance with BIPM JCGM 100:2008, Guide to the Expression of Uncertainty in Measurement (GUM), 2008, with a variable  $k$  representing a 95.45 % confidence level. Class F tolerance values are taken from NIST HB 105-1.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Note:

A positive correction indicates that the weight is heavier than the stated nominal value.  
A negative correction indicates that the weight is lighter than the stated nominal value.

Conversions:

milligram (mg) to kilogram (kg):  $\text{kg} = \text{mg} / 1000000$

milligram (mg) to gram (g):  $\text{g} = \text{mg} / 1000$

milligram (mg) to pound (lb):  $\text{lb} = \text{mg} \times 0.000002204622621848776$

milligram (mg) to ounce (oz):  $\text{oz} = \text{mg} \times 0.00003527396194958041$

Philip Lockwood  
Manager for Metrology Laboratory  
Agency Representative

Lisa Corn  
Metrologist  
Legal Signatory



**TEXAS DEPARTMENT OF AGRICULTURE**  
**COMMISSIONER SID MILLER**  
Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942  
**CALIBRATION CERTIFICATE**

**Test Completed**  
09/13/2017

**For**  
3 - Weight Kits

**Test Number**  
G-000006283

**Date Due**  
09/30/2018

**Submitted by**  
Bastrop Scale Company  
PO Drawer 2100  
Bastrop, Texas 78602

The artifacts described below have been compared to the standards of the State of Texas and were found to have the following mass corrections:

Temperature Range: 18 °C to 27 °C  
Relative Humidity Range: 40 % to 60 %  
SOP Used: NISTIR 6969, SOP No. 8, Modified Substitution

| Nominal Value | Serial / ID # | As Found Mass Correction ( mg ) | As Left Mass Correction ( mg ) | Expanded Uncertainty ( mg ) | Tolerance Class | Tolerance ( mg ) |
|---------------|---------------|---------------------------------|--------------------------------|-----------------------------|-----------------|------------------|
| 5 lb          | 6A25 A        | 72                              | 72                             | 28                          | F               | 230              |
| 5 lb          | 6A25 B        | 75                              | 75                             | 28                          | F               | 230              |
| 5 lb          | 6A25 C        | 67                              | 67                             | 28                          | F               | 230              |
| 5 lb          | 6A25 D        | 57                              | 57                             | 28                          | F               | 230              |
| 5 lb          | 6A25 E        | 71                              | 71                             | 28                          | F               | 230              |
| 5 lb          | 6A26 A        | 77                              | 77                             | 28                          | F               | 230              |
| 5 lb          | 6A26 B        | 85                              | 85                             | 28                          | F               | 230              |
| 5 lb          | 6A26 C        | 63                              | 63                             | 28                          | F               | 230              |
| 5 lb          | 6A26 D        | 74                              | 74                             | 28                          | F               | 230              |
| 5 lb          | 6A26 E        | 55                              | 55                             | 28                          | F               | 230              |
| 5 kg          | 6A27 A        | 148                             | 148                            | 59                          | F               | 500              |
| 5 kg          | 6A27 B        | 148                             | 148                            | 59                          | F               | 500              |
| 1 lb          | 6A25 A        | 6.4                             | 6.4                            | 8.4                         | F               | 70               |
| 1 lb          | 6A25 B        | 7.4                             | 7.4                            | 8.4                         | F               | 70               |
| 1 lb          | 6A25 C        | 5.4                             | 5.4                            | 8.4                         | F               | 70               |
| 1 lb          | 6A25 D        | 6.5                             | 6.5                            | 8.4                         | F               | 70               |
| 1 lb          | 6A25 E        | 20.2                            | 20.2                           | 8.4                         | F               | 70               |
| 1 lb          | 6A26 A        | 20.4                            | 20.4                           | 8.4                         | F               | 70               |
| 1 lb          | 6A26 B        | 16.9                            | 16.9                           | 8.4                         | F               | 70               |
| 1 lb          | 6A26 C        | 18.4                            | 18.4                           | 8.4                         | F               | 70               |
| 1 lb          | 6A26 D        | 17.1                            | 17.1                           | 8.4                         | F               | 70               |
| 1 lb          | 6A26 E        | 16.0                            | 16.0                           | 8.4                         | F               | 70               |
| .5 lb         | 6A25 -        | 6.5                             | 6.5                            | 5.4                         | F               | 45               |
| .5 lb         | 6A26 -        | 5.2                             | 5.2                            | 5.4                         | F               | 45               |
| 2 kg          | 6A27 A        | 50                              | 50                             | 23                          | F               | 200              |
| 2 kg          | 6A27 B        | 47                              | 47                             | 23                          | F               | 200              |
| 1 kg          | 6A27 -        | 30                              | 30                             | 12                          | F               | 100              |
| 500 g         | 6A27 A        | 15.8                            | 15.8                           | 8.4                         | F               | 70               |
| 500 g         | 6A27 B        | 24.8                            | 24.8                           | 8.4                         | F               | 70               |
| 500 g         | 6A27 C        | 16.2                            | 16.2                           | 8.4                         | F               | 70               |
| 500 g         | 6A27 D        | 25.6                            | 25.6                           | 8.4                         | F               | 70               |

❖ denotes a weight that was adjusted IAW NISTIR 6969, SOP 8.

■ denotes a weight that was rejected.

The expanded uncertainty given is in compliance with BIPM JCGM 100:2008, Guide to the Expression of Uncertainty in Measurement (GUM), 2008, with a variable *k* representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Philip Lockwood  
Manager for Metrology Laboratory  
Agency Representative

Lisa Corn  
Metrologist  
Legal Signatory



**TEXAS DEPARTMENT OF AGRICULTURE**  
**COMMISSIONER SID MILLER**  
Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942  
**CALIBRATION CERTIFICATE**

**Test Completed**  
09/13/2017

**For**  
3 - Weight Kits

**Test Number**  
G-000006283

**Date Due**  
09/30/2018

**Submitted by**  
Bastrop Scale Company  
PO Drawer 2100  
Bastrop, Texas 78602

The artifacts described below have been compared to the standards of the State of Texas and were found to have the following mass corrections:

Temperature Range: 18 °C to 27 °C  
Relative Humidity Range: 40 % to 60 %  
SOP Used: NISTIR 6969, SOP No. 8, Modified Substitution

| Nominal Value | Serial / ID # | As Found             | As Left              | Expanded         | Tolerance | Tolerance |
|---------------|---------------|----------------------|----------------------|------------------|-----------|-----------|
|               |               | Mass Correction (mg) | Mass Correction (mg) | Uncertainty (mg) | Class     | (mg)      |
| 500 g         | 6A27 E        | 15.0                 | 15.0                 | 8.4              | F         | 70        |
| .2 lb         | 6A25 -        | 6.7                  | 6.7                  | 2.1              | F         | 18        |
| .2 lb         | 6A25 *        | 3.3                  | 3.3                  | 2.1              | F         | 18        |
| .2 lb         | 6A26 -        | 4.8                  | 4.8                  | 2.1              | F         | 18        |
| .2 lb         | 6A26 *        | 5.3                  | 5.3                  | 2.1              | F         | 18        |
| .1 lb         | 6A25 -        | 2.5                  | 2.5                  | 1.1              | F         | 9.1       |
| .1 lb         | 6A26 -        | 1.2                  | 1.2                  | 1.1              | F         | 9.1       |
| .05 lb        | 6A25 -        | 1.16                 | 1.16                 | 0.53             | F         | 4.5       |
| .05 lb        | 6A26 -        | 1.16                 | 1.16                 | 0.53             | F         | 4.5       |
| 200 g         | 6A27 -        | 9.9                  | 9.9                  | 4.7              | F         | 40        |
| 200 g         | 6A27 *        | 8.7                  | 8.7                  | 4.7              | F         | 40        |
| 100 g         | 6A27 -        | 4.4                  | 4.4                  | 2.3              | F         | 20        |
| 50 g          | 6A27 -        | 4.5                  | 4.5                  | 1.2              | F         | 10        |
| .02 lb        | 6A25 -        | 0.18                 | 0.18                 | 0.21             | F         | 1.8       |
| .02 lb        | 6A25 *        | 0.63                 | 0.63                 | 0.21             | F         | 1.8       |
| .02 lb        | 6A26 -        | 0.73                 | 0.73                 | 0.21             | F         | 1.8       |
| .02 lb        | 6A26 *        | 0.19                 | 0.19                 | 0.21             | F         | 1.8       |
| .01 lb        | 6A25 -        | 0.57                 | 0.57                 | 0.18             | F         | 1.5       |
| .01 lb        | 6A26 -        | 0.57                 | 0.57                 | 0.18             | F         | 1.5       |
| .005 lb       | 6A25 -        | 0.54                 | 0.54                 | 0.15             | F         | 1.2       |
| .005 lb       | 6A26 -        | 0.60                 | 0.60                 | 0.15             | F         | 1.2       |
| .002 lb       | 6A25 -        | 0.39                 | 0.39                 | 0.10             | F         | 0.87      |
| .002 lb       | 6A25 *        | 0.23                 | 0.23                 | 0.10             | F         | 0.87      |
| .002 lb       | 6A26 -        | 0.31                 | 0.31                 | 0.10             | F         | 0.87      |
| .002 lb       | 6A26 *        | 0.17                 | 0.17                 | 0.10             | F         | 0.87      |
| .001 lb       | 6A25 -        | 0.328                | 0.328                | 0.083            | F         | 0.70      |
| .001 lb       | 6A26 -        | 0.266                | 0.266                | 0.083            | F         | 0.70      |
| 20 g          | 6A27 -        | 1.20                 | 1.20                 | 0.47             | F         | 4.0       |
| 20 g          | 6A27 *        | 1.27                 | 1.27                 | 0.47             | F         | 4.0       |
| 10 g          | 6A27 -        | 0.48                 | 0.48                 | 0.23             | F         | 2.0       |
| 5 g           | 6A27 -        | 0.42                 | 0.42                 | 0.18             | F         | 1.5       |

\* denotes a weight that was adjusted IAW NISTIR 6969, SOP 8.

■ denotes a weight that was rejected.

The expanded uncertainty given is in compliance with BIPM JCGM 100:2008, Guide to the Expression of Uncertainty in Measurement (GUM), 2008, with a variable *k* representing a 95.45 % confidence level.  
This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Philip Lockwood  
Manager for Metrology Laboratory  
Agency Representative

Lisa Corn  
Metrologist  
Legal Signatory



**TEXAS DEPARTMENT OF AGRICULTURE**  
**COMMISSIONER SID MILLER**  
 Metrology Laboratory - 1258 CR 226 / P.O. Box 1518 - Giddings, Texas 78942

**CALIBRATION CERTIFICATE**

For

3 - Weight Kits

**Test Number**  
G-000006283

**Test Completed**  
09/13/2017

**Date Due**  
09/30/2018

**Submitted by**  
Bastrop Scale Company  
PO Drawer 2100  
Bastrop, Texas 78602

The artifacts described below have been compared to the standards of the State of Texas and were found to have the following mass corrections:

Temperature Range: 18 °C to 27 °C  
 Relative Humidity Range: 40 % to 60 %  
 SOP Used: NISTIR 6969, SOP No. 8, Modified Substitution

| Nominal Value | Serial / ID # | As Found<br>Mass Correction<br>( mg ) | As Left<br>Mass Correction<br>( mg ) | Expanded<br>Uncertainty<br>( mg ) | Tolerance<br>Class | Tolerance<br>( mg ) |
|---------------|---------------|---------------------------------------|--------------------------------------|-----------------------------------|--------------------|---------------------|
| 2 g           | 6A27 -        | 0.16                                  | 0.16                                 | 0.13                              | F                  | 1.1                 |
| 2 g           | 6A27 *        | 0.32                                  | 0.32                                 | 0.13                              | F                  | 1.1                 |
| 1 g           | 6A27 -        | 0.25                                  | 0.25                                 | 0.11                              | F                  | 0.90                |

❖ denotes a weight that was adjusted IAW NISTIR 6969, SOP 8.

■ denotes a weight that was rejected.

The expanded uncertainty given is in compliance with BIPM JCGM 100:2008, Guide to the Expression of Uncertainty in Measurement (GUM), 2008, with a variable  $k$  representing a 95.45 % confidence level.

This report is not to be used to claim product endorsement by the Texas Department of Agriculture or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of the Texas Department of Agriculture Metrology Laboratory.

Philip Lockwood  
 Manager for Metrology Laboratory  
 Agency Representative

Lisa Corn  
 Metrologist  
 Legal Signatory