

CERTIFIED ORIGINAL



HOLLOWAY HOUSTON, INC.

LEADERS IN LIFTING SAFETY
World's Largest CROSBY Distributor
5833 ARMOUR DR. HOUSTON, TX 77020
PH#713-674-5631 FX#713-674-8352

Certificate of Calibration

Client: Bastrop Scale Co Inc.
Device Calibrated: 150te Compression Load Cell
Manufacturer: Coti
Model #: CG26S5
Type: Cabled Remote
Device calibrated S/N: 2011027002
Indicator(s): GSE 250 Digital Display

Invoice#: 1056341

Serial #: N/A

Calibration Units: Lbf
Date Calibrated: November 29, 2017
Verification Due: November 29, 2018
Method of calibration: Compression

Temperature: 73F*
***Correction:** n/a

Calibration Factor: N/A

This device meets the accuracy standards of ASTM E4-16(+/- 1% of reading) as indicated below.
Any other loading data is shown for reference only.

Reference True Load Lbf	As Found Reading Lbf	Run 1 Indication Lbf	Percent Error of Reading	Meets ASTM E-4 Accuracy?	Run 2 Indication Lbf	Percent Error of Reading	Meets ASTM E-4 Accuracy?	Repeatability as a percent
0		0			0			
30,000	29,850	30,250	0.83	YES	30,250	0.83	YES	0.00
50,000	49,650	50,200	0.40	YES	50,200	0.40	YES	0.00
100,000	99,100	100,200	0.20	YES	100,200	0.20	YES	0.00
150,000	148,600	150,250	0.17	YES	150,250	0.17	YES	0.00
200,000	198,050	200,150	0.08	YES	200,150	0.08	YES	0.00
250,000	247,500	250,100	0.04	YES	250,100	0.04	YES	0.00
300,000	296,900	300,000	0.00	YES	300,000	0.00	YES	0.00

[X] Adjustment made to meet specified accuracy?

HHI reference standard(s):

Serial	Calibrated	Date Due	Class A loading range	Verified by	MFG by	Temperature
B263412	11/7/2017	11/7/2019	17,500- 350,000 Lbf	NIST	Coti	23C

Calibration By: 

Date: 11-29-17

Approved By: 

Date: 11-29-17